



Facility Inspection of Laser Hair Practitioner 2023-2024

North Carolina Board of Electrolysis Examiners
Pinehurst Building, 2 Centerview Drive, Suite 60, Greensboro, NC 27407
Office: 336-856-1010 | Email: ncbeexam@att.net | www.ncbee.com

Practitioner Name:					Date of Inspection:			
Business Name:					Electrologist License #			
Physical Address:								
City/State/Zip:								
Business Phone:				E-mail:				
Employment Status: Employee ____ Sole Proprietor ____ Other _____								
Inspection Type:		Initial		New Office		Annual	Re-Inspection	
LASER DEVICES UTILIZED BY PRACTITIONER AT THIS LOCATION (use additional sheet if necessary to list all devices)								
	TYPE	MANUFACTURER	SERIAL#	CLASS	OPTICAL DENSITY	WAVELENGTH		
Device #1								
Device #2								
Device #3								
Device #4								
INSPECTION CRITERIA – for each element, enter “X” in appropriate column.							PASS	FAIL
<i>FAIL ratings in Items 14-21 will demand correction within 30 days and re-inspection at licensee’s expense.</i>								
1.	Each laser treatment room has a treatment table or other piece of furniture for placing clients for treatment.							
2.	Each laser treatment room has at least one circuline type lamp, halogen lamp, or other type of magnifying lamp.							
3.	Office has handwashing facilities on the same floor with a supply of either soap or a germicidal skin preparation for washing hands.							
4.	Office has toilet facilities in the same building with a supply of either soap or a germicidal skin preparation for washing hands.							
5.	Each laser treatment room has a supply of labeled non-sterile examination gloves.							
6.	Each laser treatment room has a supply of cotton balls and antiseptic product for cleansing clients’ skin.							
7.	Each laser treatment room has materials for cleaning instruments and other items.							
8.	Office has materials for cleaning the workplace or documentation of a cleaning contract.							
9.	Office has paper or cotton towels.							
10.	Each laser treatment room has a puncture resistant sharps container.							
11.	Office has plastic bags for containment of used materials.							
12.	Each laser treatment room has a covered trashcan.							
13.	If linens are used, a laundry bag or closed container for laundry is readily available to each treatment area.							
14.	Doors leading to each laser treatment room have laser-specific American National Standard Institute (ANSI) Z136.1 safety signs displayed.							
15.	Each laser treatment room is free of uncovered mirrors or reflective surfaces.							
16.	Laser safety eyewear is in good repair, clearly labeled with the same wavelength and optical density as the laser device operated and is worn by the client and all staff in the room, while treatment is administered.							
17.	All windows in the laser treatment area are protected from laser beam with either an opaque material or white blinds.							
18.	A fire extinguisher is readily available in each laser treatment room.							
19.	A face mask is worn by laser practitioner while laser treatment is administered.							
20.	An appropriate, functional sub-micron air filter/purifier is present and operational in each laser treatment room and powered on during treatments							
21.	A copy of the current "Supervisory Agreement" is available in the office for inspection upon request.							
22.	An up-to-date, current maintenance log exists for each laser, containing current service records as per manufacturers’ recommendation, and is available upon request.							
23.	A copy of the client’s medical history and laser treatment log along with supervising physician signed reviewed chart is available upon request.							
24.	The date of last service is prominently affixed on the laser machine itself.							
25.	An updated annual service log and calibration records for each laser is available upon request							
Inspector Comments:								

Practitioner Signature

Date

Inspector Signature

Date